THERM-*cert*™ range Product data sheet





<u>+44 (0)1946 816600</u>



brannan.co.uk

PRE-CALIBRATED JUMBO MAX MIN THERMOMETER & HYGROMETER

High accuracy thermometer pre-calibrated at two fixed temperature points and part of the Brannan **THERM-** $cert^{TM}$ range.

White moulded case with two displays for indoor temperature and humidity or outdoor temperature. Complete with 2m long cable for the outdoor probe, standard hanging hook.

Replacement probe also available.

Product functions:

- · High accuracy thermometer supplied with traceable calibration certificate
- · Maximum and minimum recording facility
- · Celsius and Fahrenheit switchable
- · 2m cable to external sensor
- Main body includes magnet, clip and stand.

Product specifications:

Temperature range: Indoor: -10 to +50°C&F

Accuracy: +/- 1°C

Outdoor: -50 to +70°C&F

Accuracy: +/-1°C between -10 and 60°C

otherwise +/-2°C

Humidity range: 20 to 90% RH

Accuracy: +/-5%RH between 30 and 90%RH

otherwise +/-7%RH

Calibration: 2 points @ 0 and 35°C

Resolution: 0.1°C / 1%RH

Battery: AAA or equivalent x 1 piece

CE/ROHS/WEEE: Compliant

Hazard information: See brannan.co.uk for information (SDS)

Product material: ABS

Thermistor probe: 2m cable with stainless steel probe

LCD dimensions: 60mm x 80mm

Product dimensions: Net: 103mm x 98mm x 21mm

Gross: 275mm x 240mm x 33mm

Product weight: Net: 121g (incl. battery & probe)

Display box weight: Packaging card: 20.5g (incl. instruction sheet)

Bubble bag: 1g

Gross: 157g (incl. certificate and packaging)

CLOCK / HUMIDIT	
Brannan England	TREMORY BE

	ISSUED BY : S.B.	rannan and Sons Lim	ited		
CALIBRATION DATE	CERTIF	CERTIFICATE NUMBER		SERIAL NUMBER	
S. Brannan and Sonds Ltd. Leconfield Industrial Estate Cleater Moor Cumbria ENGLAND CA26 SQE	Т.	Brannan hermometers and Gauges	Telephone: Fax: • Mail: sale tr www. Routered No: 2	14 (0) - 10 - 164 19460166 20 (10 (10 an) - 20 20 (12 E) (10 A)	
Instrument Type:	ELECTRONIC THE	RIMETER			
Range:	-50/70°C&F				
Resolution:	0.1°C				
Immersion:	'em	<i>u</i> ,			
Therme Building	Call the support of t	Emergent Column Temperature	Standard Instrument References		
	*c	*C			
0.1	0.0	N/A	246801/1	246801/	
34.9	35.0	N/A	246801/1	246801/	
				AA	
Signed: Shares	Calibra	tion Assistant and Approved Sig	natory		
S. Moran					
Laboratory Temperature 20*		Laboratory Humidity 201		1	

Description	Calibration points	Product no
THERM- <i>cert</i> ™ pre-calibrated jumbo	2 fixed points at	TC-006
max min thermometer and hygrometer	0 and 35°C	



