THERM-cert range Product data sheet





+44 (0)1946 816600



brannan.co.uk

PRE-CALIBRATED JUMBO MAX MIN THERMOMETER & HYGROMETER

High accuracy thermometer pre-calibrated at two fixed temperature points and part of the Brannan **THERM**-cert range.

White moulded case with two displays for indoor temperature and humidity or outdoor temperature. Complete with 2m long cable for the outdoor probe, standard hanging hook.

Replacement probe also available.

Product functions:

- · High accuracy thermometer supplied with traceable calibration certificate
- · Maximum and minimum recording facility
- · Celsius and Fahrenheit switchable
- · 2m cable to external sensor
- Main body includes magnet, clip and stand.

Product specifications:

Temperature range: Indoor: -10 to +50°C&F

Accuracy: +/- 1°C

Outdoor: -50 to +70°C&F

Accuracy: +/-1°C between -10 and 60°C

otherwise +/-2°C

Humidity range: 20 to 90% RH

Accuracy: +/-5%RH between 30 and 90%RH

otherwise +/-7%RH

Calibration: 2 points @ 0 and 35°C

Resolution: 0.1°C / 1%RH

Battery: AAA or equivalent x 1 piece

CE/ROHS/WEEE: Compliant

Hazard information: See brannan.co.uk for information (SDS)

Product material: ABS

Thermistor probe: 2m cable with stainless steel probe

LCD dimensions: 60mm x 80mm

Product dimensions: Net: 103mm x 98mm x 21mm

Gross: 275mm x 240mm x 33mm Net: 121g (incl. battery & probe)

Product weight: Net: 121g (incl. battery & probe)

Display box weight: Packaging card: 20.5g (incl. instruction sheet)

Bubble bag: 1g

Gross: 157g (incl. certificate and packaging)

CLOCK / HUMIDIT	
Brannan England	TREMORY BE

		BRATION C		CAII	
		annan and Sons Lim			
CALIBRATION DATE	CERTIF	CERTIFICATE NUMBER		SERIAL NUMBE	
S. Brannan and Sonds Ltd. Leconfield Industrial Estate Cleater Moor Cumbria ENGLAND CA26 SQE	The state of the s	Grannan hermometers and Gauges	Telephone: Fax: - Mail: sale tr www. Recurred No: 2	w.b	
Instrument Type:	ELECTRONIC THER	RETER			
Range:	-50/70°C8F				
Resolution:	0.1°C	111			
Immersion:	gum	<i>u</i> ,			
Therme	California Simperato	Emergent Column Temperature	Standard Instrument References		
	*C	*C			
0.1	0.0	N/A	246801/1	246801/	
34.9	35.0	N/A	246801/1	246801	
			3	WES S	
Signed: S. Moran	Calibrat	ion Assistant and Approved Sig	natory		
g. moren			100	120	

Description	Calibration points	Product no
THERM- <i>cert</i> pre-calibrated jumbo max min thermometer and hygrometer	2 fixed points at 0 and 35°C	TC-006
max min thermometer and mygrometer	U and 33 U	



