

THERM-cert™ range

Product data sheet

✉ sales@brannan.co.uk

☎ +44 (0)1946 816600

🌐 brannan.co.uk

PRE-CALIBRATED 215MM PLASTIC WALL WORKPLACE THERMOMETER

215mm wall thermometer pre-calibrated as part of the Brannan **THERM-cert™** range. Plastic wall thermometer with recommended temperature for workplace in accordance with The Workplace (Health, Safety & Welfare) Regulations 1992 Section 7).

Product features:

- Supplied with traceable calibration certificate
- 215mm, highly visible wall thermometer
- Recommended temperature highlighted
- Made in UK

Product specifications:

Range:	-20 to +50°C&F*
Divisions:	1°C/F
Accuracy:	+/- 1.0°C @ 10°C
CE/ROHS/WEEE:	N/A
Hazard information (SDS):	See brannan.co.uk for information
Product material:	Moulding: ABS Glass tube: kerosene filled
Component weights:	Plastic - 16g Glass - 6g
Product weight:	Unpackaged - 22g Gross - 58.5g (incl. certificate & packaging)
Product dimensions:	Unpackaged - 215mm x 40mm x 8mm Packaged - 340mm x 250mm x 20mm



Description	Calibration points	Product no
THERM-cert™ pre-calibrated 215mm plastic workplace thermometer	1 fixed points at 16°C	TC-017

TRACEABLE CALIBRATION CERTIFICATE

ISSUED BY : S Brannan and Sons Limited

CALIBRATION DATE	CERTIFICATE NUMBER	SERIAL NUMBER
<p>S. Brannan and Sons Ltd. Leconfield Industrial Estate Custer Road SHROPSHIRE GUAR 8GE</p> <p>Brannan Thermometers and Gasps</p> <p>Telephone: 01946 816600 Fax: 01946 816601 Email: sales@brannan.co.uk Website: www.brannan.co.uk Registered No: 20613 ENGLAND</p>		
<p>Instrument Type: 215MM WORKPLACE THERMOMETER Range: -25/+50°C&F</p>		
Thermometer Reading 15.5	Lab Temperature 16.0	Emergent Column Temperature N/A
		Standard Instrument References 246801/1 246801/2
<p>Signed:  Calibration Assistant and Approved Signatory</p> <p>S. Meen Laboratory Temperature 20°C +/- 5°C Laboratory Humidity 20% to 90%</p>		

S. Brannan and Sons operate an ISO9001 certified Quality Management System in order to ensure the accuracy and reliability of every instrument we produce.
The readings were determined by comparison with traceable standard instruments, the reference numbers of which are shown above.
We recommend that periodic recalibration should be carried out on this instrument yearly, but the recalibration frequency to best determined by the user depending upon the frequency and conditions of use, as well as the level of accuracy required by the process.

* PLEASE NOTE: Scale ranges may differ slightly from those advertised due to manufacturing procedures