

THERM-cert range

Product data sheet

✉ sales@brannan.co.uk

☎ +44 (0)1946 816600

🌐 brannan.co.uk

PRE-CALIBRATED TWIN READING THERMOMETER

High accuracy thermometer pre-calibrated at two fixed temperature points and part of the Brannan **THERM-cert** range.
Electronic thermometer with max min feature and high/low temperature alarm setting. Ideally suited to read temperature in a fridge or freezer or outdoor temperature and room temperature simultaneously.

Product features:

- High accuracy thermometer supplied with traceable calibration certificate
- Choice of calibration points to suit various applications
- Max min recording facility
- Celsius and Fahrenheit switchable
- 2m cable to external sensor
- Magnetic fixture on back of meter



Product specifications:

Internal temperature range:	-20 to +50°C&F
External temperature range:	-50 to +70°C&F
Accuracy:	Internal: +/- 1°C External: +/- 1.5% between -50 & -20°C +/- 1.0% between -20 & +70°C
Calibration:	2 points @ 0 and 35°C
Resolution:	0.1°C
Sampling time:	10 seconds
Battery:	1 x AAA (supplied)
CE & RoHS/WEEE:	Compliant
Hazard information (SDS):	See brannan.co.uk for information
Probe lengths:	21mm with 2m cable
LCD dimensions:	21mm x 47mm
Dimensions:	Net: 63mm x 63mm x 15mm Gross: 275mm x 240mm x 39mm
Weights:	Meter: 62g (including battery) Display box: 22.5g Gross: 100.5g (incl. certificate and packaging)

TRACEABLE CALIBRATION CERTIFICATE			
ISSUED BY: S.Brannan and Sons Limited			
CALIBRATION DATE	CERTIFICATE NUMBER	SERIAL NUMBER	
<div> <div> S. Brannan and Sons Ltd Lancaster Industrial Estate Culter Moor Cumbria ENGLAND CA25 9QE </div> <div> Thermometers and Gauges </div> <div> Telephone: +44 (0) 1946 816600 Fax: +44 (0) 1946 816625 E Mail: sales@brannan.co.uk Web: www.brannan.co.uk Registered No: 290512 ENGLAND </div> </div>			
Instrument Type: ELECTRONIC THERMOMETER			
Range: -50/70°C&F			
Resolution: 0.1°C			
Immersion: 76mm			
Thermometer Reading	Calibration Temperature	Environ Column Temperature	Standard Instrument References
0.1	0.1	N/A	246801/1 246801/2
5.1	35.0	N/A	246801/1 246801/2
<div> <div>Signed: S. Moran</div> <div>Calibration Assistant and Approved Signatory</div> </div>			
Laboratory Temperature 20°C +/- 5°C Laboratory Humidity 20% to 90%			
<small> S.Brannan and Sons operate an ISO9001 certified Quality Management System in order to ensure the accuracy and reliability of every instrument we produce. The readings were determined by comparison with traceable standard instruments, the reference numbers of which are shown above. We recommend that periodic recalibration should be carried out on this instrument yearly, but the recalibration frequency is best determined by the user depending upon the frequency and conditions of use, as well as the level of accuracy required by the process. </small>			

Description	Calibration points	Product no
THERM-cert pre-calibrated twin reading thermometer	fixed at 0 and 35°C	TC-007
	fixed at -18 and 0°C	TC-012



Measurement for life.

